

Patient Commission/Configuration ID: _____

Functional Unit Serial Number: _____ Date of Invoice: _____

1. Maintenance after 18 Months* (from date of invoice)

Item	20mm	Description	Potential Problem	Measure	Done	Date
3	see main-tenance plan	sliding washer, Ø = 26mm	wear	replacing		
7	VE3771-012/26	O-ring damper**	wear	replacing		
8	GS1108-500	sliding bushing**	wear	replacing		
9	FE1027-01	coil spring**	wear	replacing		
15	VE3771-11/10	O-ring for securing the spring unit	wear	replacing		
-	SH5805-15/18	spring unit, blue, normal, max. 15° range of motion	wear radial move of disc springs noise of spring unit	replacing		
-	SH5805-15/25	spring unit, green, medium, max. 15° range of motion		realigning disc springs with pliers		
-	SH5805-10/40	spring unit, white, strong, max. 10° range of motion		greasing spring unit with spray oil (article no. FT3000-15)		
-	SH5805-10/60	spring unit, yellow, very strong, max. 10° range of motion				
-	SH5805-05/99	spring unit, red, extra strong, max. 5° range of motion				
18	BR1211-L030	sliding bushing	wear	replacing		

* depending on the assessment of the distributor of the custom-made product regarding the patient's usage behaviour

** is part of the functional unit

Location, Date _____

Signature of the Orthotist _____

2. Maintenance after 36 Months* (from date of invoice)

Item	20mm	Description	Potential Problem	Measure	Done	Date
1	SB1069-L0960	bearing nut	wear	replacing		
3	see main-tenance plan	sliding washer, Ø = 26mm	wear	replacing		
5	SC1403-L10	countersunk flat head screw with hexalobular socket	wear	replacing		
6	SH0493-01	plunger**	wear	replacing		
7	VE3771-012/26	O-ring damper**	wear	replacing		
8	GS1108-500	sliding bushing**	wear	replacing		

PR9152-GB-2025-08

Patient Commission/Configuration ID: _____

Functional Unit Serial Number: _____ Date of Invoice: _____

2. Maintenance after 36 Months* (from date of invoice)

Item	20mm	Description	Potential Problem	Measure	Done	Date
9	FE1027-01	coil spring**	wear	replacing		
10	SC1403-L08/1	countersunk flat head screw with hexalobular socket	wear	replacing		
13	SC1416-L14	countersunk flat head screw with hexalobular socket**	wear	replacing		
14	SC1416-L14	countersunk flat head screw with hexalobular socket (axle screw)**				
4-14	SH8975-AL	functional unit	loss of function, wear	replacing		
15	VE3771-11/10	O-ring for securing the spring unit	wear	replacing		
-	SH5805-15/18	spring unit, blue, normal, max. 15° range of motion	wear radial move of disc springs noise of spring unit	replacing		
-	SH5805-15/25	spring unit, green, medium, max. 15° range of motion		realigning disc springs with pliers		
-	SH5805-10/40	spring unit, white, strong, max. 10° range of motion		greasing coil spring with spray oil (article no. FT3000-15)		
-	SH5805-10/60	spring unit, yellow, very strong, max. 10° range of motion				
-	SH5805-05/99	spring unit, red, extra strong, max. 5° range of motion				
16	ET3860	controller with Bluetooth®*** with lithium-polymer battery	outdated software	updating software		
			bad battery health	replacing		
18	BR1211-L030	sliding bushing	wear	replacing		

* depending on the assessment of the distributor of the custom-made product regarding the patient's usage behaviour

** is part of the functional unit

*** The Bluetooth® word mark and logos are registered trademarks of Bluetooth SIG, Inc. and any use of such marks by FIOR & GENTZ is under license.

Location, Date _____

Signature of the Orthotist _____